

**UNIVERSITI TEKNOLOGI MARA**



**THE EFFECT OF LEAD SHIELDING THICKNESS  
IN COMPUTED TOMOGRAPHY OF THE  
HEAD: A PHANTOM STUDY**

**FARIDA AIMI BINTI MUSTAPHA**

**Dissertation submitted in partial fulfillment of the requirements for the  
degree of Bachelor of Medical Imaging (Hons.)**

**Faculty of Health Sciences**

**June 2015**

## AUTHOR'S DECLARATION

I declare that the in tis thesis was carried out in accordance with the regulations of Unniversiti Teknologi MARA. It is original and is the results of my own work, unless otherwise indicated or acknowledge as referenced work. This thesis has not been submitted to any other academic institution or non-academic institution for any degree or qualification.

I, hereby, acknowledge that I have been supplied with the Academic Rules and Regulations for Under Graduate, Universiti Teknologi MARA, regulating the conduct of my study and research.

Nama of Student	:	Farida Aimi binti Mustapha
Student I.D. No.	:	2011485202
Programme	:	Bachelor of Medical Imaging – HS222
Faculty	:	Health Science
Thesis Title	:	The effect of lead shielding thickness in Computed Tomography of the head: A phantom Study
Signature of Student	:	.....
Date	:	June 2015

## ABSTRACT

CT of the head is the most requested examination in CT scan. However, the routine use of external shielding in CT imaging appears to be limited. It is because of concern that the application of external lead shielding may result in increased internal scatter and consequently lead to an increased dose to the radiosensitive organs (Palmer, 2008). The purpose of this study is to investigate the effect of different thickness of lead shielding in computed tomography (CT) examination of the head. This research is an experimental study using an anthropomorphic Rando phantom to imitate the patient. Different thickness of thyroid shield was applied which are 0.0 mm Pb, 0.5 mm Pb, 1.0 mm Pb and 1.5 mm Pb to protect thyroid gland from scattered radiation. Four thermoluminescent detector (TLD) chips were placed at surface of thyroid gland and 1 cm depth within thyroid tissue upon slice 10 and slice 11 of phantom. The Rando phantom was placed on the Siemens Somatom Definition Flash Dual Source 128-slice CT scanner table and a CT head protocol was performed. Then TLDs were read to measure the surface dose and absorbed dose of thyroid gland using different thickness of lead shield. The data collected was analysed to measure the significant different radiation dose with various different thickness of lead shielding using statistical one way ANOVA. This study provides evidence that the effectiveness of lead shielding varied with their thickness respectively corresponding to thyroid dose due to scattered radiation. The highest dose reduction was achieved by 1.5 mm Pb thickness which is 57.48% at the surface, 55.77% at 1 cm depth within thyroid tissue, followed by 1.0mmPb result in 56.69% at surface, 50% at 1 cm depth. The lowest dose reduction was achieved with 0.5 mm Pb is 52.76% at surface, 41.35% at 1 cm depth within thyroid tissue. This research suggested the application of thyroid shielding gives advantage especially to the paediatric patient and patients that require repetition CT examination because long term illness. This research has proven that reducing radiation dose to the patient because of scattered radiation by increasing the thickness of lead shielding can further decrease dose to radiosensitive organ outside the scanning area. As a result, the risks of inducing cancer will also be decreased.

# TABLE OF CONTENTS

<b>AUTHOR'S DECLARATION</b>	<b>ii</b>
<b>ABSTRACT</b>	<b>iii</b>
<b>ACKNOWLEDGEMENT</b>	<b>iv</b>
<b>LIST OF TABLES</b>	<b>viii</b>
<b>LIST OF FIGURES</b>	<b>ix</b>
<b>LIST OF SYMBOLS</b>	<b>x</b>
<b>LIST OF ABBREVIATIONS / NOMENCLATURE</b>	<b>xi</b>
<b>CHAPTER ONE: INTRODUCTION</b>	<b>1</b>
1.1 Research Background	1
1.2 Problem Statement	2
1.3 Rational Of Study	4
1.4 Objectives	5
1.4.1 General Objective	5
1.4.2 Specific Objectives	5
1.5 Hypothesis	5
1.5.1 Hypothesis 1	5
1.5.2 Hypothesis 2	5
1.5.3 Hypothesis 3	5
1.5.4 Hypothesis 4	5
<b>CHAPTER 2: LITERATURE REVIEW</b>	<b>6</b>
2.1 Advantages Of Computed Tomography (CT) Over Conventional Radiography	6
2.2 Computed Tomography (CT) Of The Head	7
2.3 Radiation Dose In Computed Tomography (CT) Scan	7
2.4 Shielding As Dose Reduction in CT Of Head	9

<b>CHAPTER THREE: METHODOLOGY</b>	<b>10</b>
3.1 Research Design	10
3.3 Research Setting	12
3.4 Sample Design And Sample Size	12
3.4.1 Inclusion Criteria And Exclusion Criteria	12
3.4.1.1 Inclusion Criteria	12
3.4.1.2 Exclusion Criteria	12
3.5 Data Collection Procedure	13
3.5.1 Equipment	13
3.5.1.1 The 128 –Slice Computed Tomography (CT)	13
3.5.1.2 The Anthropomorphic Head Phantom (RANDO® Phantoms)	13
3.5.1.3 The 0.5 mm Pb Thyroid Shields	14
3.5.1.4 Thermoluminescent Dosimeter (TLD) Chips	14
3.5.1.5 Harshaw 3500 TLD Reader	14
3.5.1.6 Radcal Accu-Pro	15
3.5.2 Imaging Procedure for Dose Measurements.	15
3.5.2.1 TLD Measurement	15
3.5.3 Scanning Procedure And Protocol	16
3.5.4 Calculation	17
3.5.5 Statistical Analysis	18
3.5.6 Ethical Consideration	19
<b>CHAPTER 4: RESULT</b>	<b>21</b>
4.1 Descriptive Statistic	21
4.1.1 Mean Value of Surface Dose for Different Thickness of Lead Shielding.	21
4.1.2 Mean Value of Absorbed Dose for Different Thickness of Lead Shielding.	23
4.1.3 Mean Value of both Surface And Absorbed Dose for Different Thickness of Lead Shielding.	25
4.2 One Way Anova	26