

UNIVERSITI TEKNOLOGI MARA

**FINGER EXOSKELETON FOR
EARLY ACUTE STROKE
REHABILITATION: CONTROL
DESIGN AND PERFORMANCE
ANALYSIS**

MUHAMMAD NAIM BIN ABDU SALAM

Dissertation submitted in partial fulfillment
of the requirements for the degree of
Master of Science
(Mechanical Engineering)

Faculty of Mechanical Engineering

March 2020

AUTHOR'S DECLARATION

I declare that the work in this dissertation was carried out in accordance with the regulations of Universiti Teknologi MARA. It is original and is the results of my own work, unless otherwise indicated or acknowledged as referenced work. This thesis has not been submitted to any other academic institution or non-academic institution for any degree or qualification.

I, hereby, acknowledge that I have been supplied with the Academic Rules and Regulations for Post Graduate, Universiti Teknologi MARA, regulating the conduct of my study and research.

Name of Student : Muhammad Naim Bin Abdu Salam

Student I.D. No. : 2017986763

Programme : Master of Science (Mechanical Engineering) –
EM703

Faculty : Mechanical Engineering

Dissertation Title : Finger Exoskeleton for Early Acute Stroke
Rehabilitation: Control Design and Performance
Analysis

Signature of Student :

Date : March 2020

ABSTRACT

The number of stroke patients suffering from upper extremity hemiparesis increases from year to year. Manual repetitive therapy applied to stroke patients by trained therapists is limited by the number of therapists available. Many types of rehabilitation devices are invented to help patients recover their hand functions including exoskeleton devices. These devices can deliver high intensity therapy for a longer period of time. Every exoskeleton rehabilitation device is made up of three components: exoskeleton mechanism, actuator and control system. Control system is central to rehabilitation devices. The accurate control algorithm implementation is very crucial for the rehabilitation aspect. The performance of the control algorithm for rehabilitation must be investigated to assess the efficiency of the devices for rehabilitation. This research aimed to apply different Proportional Integral and Derivative (PID) controller configurations for position control of the index finger exoskeleton in rehabilitation. The second objective was to assess the performance of position control for index finger in rehabilitation. The control algorithm utilized was PID control algorithm. The parameter controlled was position of index finger measured through Metacarpophalangeal (MCP) joint angle. PID parameters were tuned using heuristic method and Ziegler Nichols' closed loop method. In order to assess the performance of the control system, desired or reference values were used. The actual data from position control were compared to reference values to investigate the performance. This study suggested that the PID controller managed to control the parameter according to the set point better than Proportional (P), Proportional and Integral (PI), and Proportional and Derivative (PD) controllers. The device could be manipulated to move to any position within the range of exoskeleton motion accurately. The results of the study also suggested that the performance of the control system for rehabilitation was very good as the error (difference between actual value and reference value) was less than 10%. Hence, the objectives of the research project were successfully achieved.

ACKNOWLEDGEMENT

Firstly, I wish to thank God for giving me the opportunity to embark on my Master and for completing this long and challenging journey successfully. My gratitude and thanks go to my supervisor Dr. Mohd Nor Azmi bin Ab Patar, and co-supervisor, Prof. Ir. Dr. Jamaluddin Mahmud. Thank you for the support, patience and ideas in assisting me with this project.

Finally, this thesis is dedicated to my beloved father and mother for their vision and determination to educate me. This piece of victory is dedicated to both of you. Alhamdulillah.

TABLE OF CONTENTS

	Page
CONFIRMATION BY PANEL OF EXAMINERS	ii
AUTHOR'S DECLARATION	iii
ABSTRACT	iv
ACKNOWLEDGEMENT	v
TABLE OF CONTENTS	vi
LIST OF TABLES	viii
LIST OF FIGURES	ix
LIST OF PLATES	xi
LIST OF SYMBOLS	xii
LIST OF ABBREVIATIONS	xiii
LIST OF NOMENCLATURE	xiv
CHAPTER ONE INTRODUCTION	1
1.1 Research Background	1
1.2 Problem Statement	2
1.3 Objectives	3
1.4 Scope of Work	3
1.5 Significance of Study	4
CHAPTER TWO LITERATURE REVIEW	5
2.1 Introduction	5
2.1.1 Stroke Recovery Stages	6
2.1.2 Robotic Hand Exoskeleton for Stroke Rehabilitation	8
2.2 Previous Studies on Hand Exoskeleton	10
2.3 Multivariate Linear Regression	13
2.3.1 Hand Exoskeleton Mechanism	14
2.3.2 Actuator Mechanism	16
2.3.3 Control Strategy	17
2.4 Position Control	19